

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130132 ✓
Name	Matthew E Armstrong	ID#	5906 ✓
		Date	October 29, 2010 ✓
A	Agency	Seward Police Department	Phone # (907) 224-3338
Instrument Location 410 Adams St Seward Community Jail, Seward AK			
B	Alco S/N	G30459 ✓	Target Value .078 .075 ✓
			High Pressure 700
	Alco Test Values	.075 ✓ 1st Alco	.073 ✓ 2nd Alco
Signature <u>Matthew E Armstrong</u>			

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of DEC, 2010.

Daniel L. Nicholas (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

OCTOBER 29, 2010

OPERATOR'S NAME:

ARMSTRONG, MATTHEW E

OPERATOR'S NUMBER: 5906

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 0000000

DEPT/AGENCY: SMD1

CASE/REPORT: 10-0000

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: 630459

— BREATH ANALYSIS —

.078 ADJUSTED FOR 29.07 in
ALCO TARGET .075 09:07
BLANK TEST .000 09:08
INTERNAL STANDARD VERIFIED 09:08
ALCO TV 29.07 in .075 09:08
BLANK TEST .000 09:09
SUBJECT SAMPLE .000 09:10
BLANK TEST .000 09:10
ALCO TV 29.07 in .073 09:11
BLANK TEST .000 09:11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

OCTOBER 29, 2010

TIME 09:04

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHARGER: 40c

BREATH TUBE: 44c

BAROMETER: 29.07 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~